



Republic of the Philippines
Cagayan State University
BIDS & AWARDS COMMITTEE

Tuguegarao City, Cagayan
Tel. No.: (078) 377-5520 website: www.csu.edu.ph

BID BULLETIN NO: 2023-08-02

August 11, 2023

This **BID BULLETIN NO: 2023-08-02: CLARIFICATION** is issued to direct all prospective bidders to observe the changes/amendment/clarification for the project: **SUPPLY AND DELIVERY OF CAHS LABORATORY EQUIPMENT (PROJECT REF.NO: CSU2023-08-01-CAHS-LAB-EQUIP-ANDREWS)**.

ORIGINAL POSTING	New / Amended
THERMOCYCLER <ul style="list-style-type: none"> ➤ Temperature range of well: 3°C to 100 °C 	THERMOCYCLER <ul style="list-style-type: none"> ➤ Temperature range of well: 4°C to 99.9 °C or better
GEL ELECTROPHORESIS MACHINE <ul style="list-style-type: none"> ➤ Can run multiple gels in one run 	HORIZONTAL GEL ELECTROPHORESIS MACHINE <ul style="list-style-type: none"> ➤ Can run 2 gels at once ➤ Smallest gel size should hold minimum of 8 wells
MICROWAVE MINIMUM AND BASIC TECHNICAL SPECIFICATIONS: <ul style="list-style-type: none"> ➤ Power: 220V ➤ With digital display ➤ Two-year warranty on parts and service 	MICROWAVE MINIMUM AND BASIC TECHNICAL SPECIFICATIONS: <ul style="list-style-type: none"> ➤ Power: 220V ➤ With digital display ➤ Two-year warranty on parts and service ➤ Capacity: Min. 25L
GEL DOCUMENTATION SYSTEM <ul style="list-style-type: none"> ➤ Comes with face shield for protection against UV when cutting bands 	GEL DOCUMENTATION SYSTEM <ul style="list-style-type: none"> ➤ Comes with accessories to protect the user from UV rays when cutting out bands
REFRIGERATED CENTRIFUGE <ul style="list-style-type: none"> ➤ Includes a swing-out rotor for tubes with adapters with lids for 15 mL and 50 mL tubes. 	REFRIGERATED CENTRIFUGE <ul style="list-style-type: none"> ➤ Includes a swing-out rotor for conical tubes with adapters with lids for 15ml and 50ml conical tubes
MULTICHANNEL PIPETTOR <ul style="list-style-type: none"> ➤ Set contains a 0.5 µL to 20 µL volume and a 30-300 µL volume pipette 	MULTICHANNEL PIPETTOR <ul style="list-style-type: none"> ➤ 8 Channels ➤ Set contains a 2 µL to 20 µL (minimum volume can be lower) and a 30-300 µL volume pipette



Republic of the Philippines
Cagayan State University
BIDS & AWARDS COMMITTEE

Tuguegarao City, Cagayan
 Tel. No.: (078) 377-5520 website: www.csu.edu.ph

<p>SET OF PIPETTORS WITH STAND</p> <ul style="list-style-type: none"> ➤ 1 set must include: <ul style="list-style-type: none"> ○ 0.2 μL to 2.5 μL ○ 0.5 μL to 10 μL ○ 1 μL to 10 μL ○ 10 μL to 100 μL ○ 100 μL to 1000 μL 	<p>SET OF PIPETTORS WITH STAND</p> <ul style="list-style-type: none"> ➤ 1 set must include: <ul style="list-style-type: none"> ○ 0.2 μL to 2.5 μL ○ 1 μL to 10 μL ○ 10 μL to 100 μL ○ 100 μL to 1000 μL
<p>Delivery Period</p> <p>90 Calendar days</p>	<p>120 Calendar days</p>

Note:

The Committee shall observe the conditions provided for in Sections 23.4.1.3 of the 206 R-IRR, RA 9184, within the relevant period as provided in the Bidding Documents in the case of Goods.

For guidance and information of all concerned.

In case of inquiry, please contact BAC Secretariat at (078) 377-5520 and bacsecretariat@csu.edu.ph.

MR. ABRAHAM C. BANGAYAN, JD, MLS
 Chairperson
 Bids & Awards Committee - 2